

This example shows you how to acquire data using National Instruments USB-6009 multifunction I/O devices. It will show how to plot the acquired data dynamically both in Origin and LabView.

Notes: This VI is created with LabView 7.1.

Make sure that you have installed NA-DAQmx in order to open these DAQ related VIs, and the Labview version you're running is supported by the installed version of NA\_DAOmx. e.g. For Labview 2012, it is required to first install NA-DAQmx 9.6.

For detailed information, please refer to the website of National Instruments.

The following is the diagram of this VI.

- 

